



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2024/B/5231909  
Dated/दिनांक : 31-07-2024

### Bid Document/ बिड दस्तावेज़

Bid Details/बिड विवरण	
Bid End Date/Time/बिड बंद होने की तारीख/समय	21-08-2024 16:00:00
Bid Opening Date/Time/बिड खुलने की तारीख/समय	21-08-2024 16:30:00
Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)	90 (Days)
Ministry/State Name/मंत्रालय/राज्य का नाम	Ministry Of Science And Technology
Department Name/विभाग का नाम	Department Of Scientific And Industrial Research (dsir)
Organisation Name/संगठन का नाम	Council Of Scientific And Industrial Research (csir)
Office Name/कार्यालय का नाम	Central Drug Research Institute
Total Quantity/कुल मात्रा	1
Item Category/मद केटेगरी	Gas Chromatography (Q3)
MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है	Yes
Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है	Yes
Document required from seller/विक्रेता से मांगे गए दस्तावेज़	Experience Criteria,Certificate (Requested in ATC),OEM Authorization Certificate,Additional Doc 1 (Requested in ATC) *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया	No
Type of Bid/बिड का प्रकार	Two Packet Bid
Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय	3 Days
Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
Evaluation Method/मूल्यांकन पद्धति	Total value wise evaluation

**EMD Detail/ईएमडी विवरण**

Required/आवश्यकता	No
-------------------	----

**ePBG Detail/ईपीबीजी विवरण**

Advisory Bank/एडवाइजरी बैंक	State Bank of India
ePBG Percentage(%) / ईपीबीजी प्रतिशत (%)	3.00
Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने).	26

(a). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

**Beneficiary/लाभार्थी :**

DIRECTOR

Central Drug Research Institute, Department of Scientific and Industrial Research (DSIR), Council of Scientific and Industrial Research (CSIR), Ministry of Science and Technology, Lucknow - 226031  
(The Director Csir-cdri)

**Splitting/विभाजन**

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

**MII Purchase Preference/एमआईआई खरीद वरीयता**

MII Purchase Preference/एमआईआई खरीद वरीयता	Yes
--	-----

**MSE Purchase Preference/एमएसई खरीद वरीयता**

MSE Purchase Preference/एमएसई खरीद वरीयता	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

3. Preference to Make In India products (For bids > 200 Crore) (can also be used in Bids < 200 Crore but only after exemption by competent authority as defined in Deptt of Expenditure OM dated 28.5.2020): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

5. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

## Gas Chromatography ( 1 pieces )

**(Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)**

### Technical Specifications/तकनीकी विशिष्टियाँ

\* As per GeM Category Specification/जेम कैटेगरी विशिष्टि के अनुसार

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
GAS CHROMATOGRAPH (GC) CONFIGURATION	"GC Detectors inclusive in the scope of supply (Select Only those Detectors, which are part of supply)"	Flame Ionization Detector (FID)
	"Number of FID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	2

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	"Number of TCD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	"Number of ECD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	"Number of NPD/FTD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	"Number of FPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	"Number of PID in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	"Number of PFPD in the System - inclusive in the scope of Supply (Select '0' if this Detector is not a part of supply)"	0
	Number of Inlets GC can Simultaneously Support	2
	Number of Detectors (Ports) GC can Simultaneously Support	2
	Number of Capillary Columns provided in GC - inclusive in the scope of Supply	1
	Type of Capillary Column	Polar
	"GC Capillary Column Length (Hint:-Select '0' if Not Applicable)"	60.0 (meter)

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	"Number of Packed Columns provided in GC - inclusive in the scope of Supply (Hint:-Select '0' if Not Applicable)"	2
	"GC Packed Column Length (Hint:-Select '0' if Not Applicable)"	6.0 (foot)
PERFORMANCE	"GC Detector Minimum Detection Level (MDL) "	< 1.4 pg C/s for FID, < 1.8 pg C/s for FID
COLUMN OVEN (GAS CHROMATOGRAPH - GC)	Maximum Operating Temperature (GC Oven)	upto 450 deg. C (2) Or higher
	Number of Oven Ramps / Plateaus for Temperature programming	6/7 or more
	Temperature Ramp Rate - Maximum	120 Deg. C/minute or better
PNEUMATIC CONTROL	Number of Injection Ports	1
	Maximum Injection Port heating Temperature	upto 400 deg. C (1), upto 450 deg. C (2) Or higher
	Pressure Range	upto100 psi (1), upto 120 psi (2) Or higher
	Pressure set point Resolution ( $\pm$ )(psi)	0.001 Or lower
	Type of Sample Injection Port	split / Splitless
	Split Ratio	7500:1
AUTO INJECTOR / AUTO SAMPLER	Compatible Liquid Auto Sampler - inclusive in the scope of supply	Yes
	Capacity of the Liquid Auto Sampler (Number of Vials)	0, 15, 100, 150 Or higher
	Number of Injection Ports in the Liquid Auto Sampler	0
	Number of Channels in the Liquid Auto Sampler	0
	Minimum Volume of the injected liquid sample	1.0 (microliter)
	Maximum Volume of the injected liquid sample	10.0 (microliter)
	"Liquid Auto Sampler Features (Select applicable features Only)"	Fully automated Liquid Auto Sampler, Injects with variable speed & varying syringe sizes, Operates through Software (Software inclusive in the scope of supply), NA

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
HEAD SPACE SAMPLER	<b>Compatible Head Space Sampler- inclusive in the scope of supply</b>	Yes
	<b>Capacity of the Head Space Sampler Carousel (number of Vials)</b>	12 or better
	<b>Head Space Sampler Temperature Range</b>	Upto 200 deg. C or better
PYROLYZER SYSTEM	<b>Compatible Pyrolyzer System for pyrolysis of solid and liquid samples- inclusive in the scope of supply</b>	Yes
SALIENT FEATURES OF GAS CHROMATOGRAPH (GC)	<b>"Salient Features of GC (Select applicable features Only)"</b>	Automatic leak testing/detection of Pneumatics
SOFTWARE	<b>"Software Features (Select Applicable Features only)"</b>	Compatible with Windows 7 or higher version operating system
GENERIC	<b>Display</b>	Through PC (PC is not part of supply), Built-in Display
	<b>Connectivity interface</b>	USB, ETHERNET, RS 232
	<b>Provision to split the sample from single injection to two different detectors and get results for two detectors simultaneously</b>	Yes
ACCESSORIES	<b>Gas Supply inclusive in the scope of supply</b>	Nitrogen, Hydrogen, Zero Air
	<b>Capacity of Gas cylinder</b>	47 liters water capacity
	<b>Number of Cylinders of each type supplied - inclusive in the scope of supply</b>	1
	<b>Gas cylinders certified by PESO</b>	Yes
	<b>Regulator - inclusive in the scope of supply</b>	Dual Stage Regulator
	<b>"Other Accessories inclusive in the scope of supply (Select only those Accessories which are part of supply)"</b>	Gas purification panel for all Gases with moisture trap, Hydrocarbon trap & Oxygen trap, Syringe for manual injection x 05 nos, Column Cutter x 2 nos
ENVIRONMENTAL	<b>Minimum Operating Temperature</b>	0.0 - 10.0 Or lower <b>(degree Celsius)</b>
	<b>Maximum Operating Temperature</b>	65.0 - 65.0 Or higher <b>(degree Celsius)</b>

Specification	Specification Name/विशिष्टि का नाम	Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य
	<b>Operating Humidity (RH) (%) at 40 degree C</b>	95.0 - 95.0 Or higher ( <b>percent</b> )
INSTALLATION & COMMISSIONING / WARRANTY / AVAILABILITY OF SPARES / TRAINING	<b>Warranty on equipment</b>	2.0 ( <b>year</b> )
	<b>Number of days onsite training provided</b>	10.0 ( <b>day</b> )
CERTIFICATION / TEST REPORT	Certification	ISO, CE
	"Availability of UL/CE Certificate as per IEC 61010- 1 latest - Safety requirements for electrical equipment for measurement,control, and laboratory use General requirements"	Yes, No
	<b>Conformance to EMC/EMI as per IEC 61326-1 latest - Electrical equipment for measurement, control and laboratory use</b>	No
	<b>Availability of Test Reports from Central Govt / NABL approved / ILAC accredited Lab to prove conformity to the specification</b>	Yes, No

#### Additional Specification Parameters - Gas Chromatography ( 1 pieces )

Specification Parameter Name	Bid Requirement (Allowed Values)
Specification	Bidder must comply the specification attached in scope of supply and Text Based ATC. The Technical evaluation shall be done on the basis of text based ATC
Eligibility criteria	Bidder must submit Bid Securing Declaration as per attached format attached in Bid document (if not submit this declaration alongwith bid. The bid will be summarily rejected)

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

S.No./क्र. सं.	Consignee Reporting/Officer/ परेषिती/रिपोर्टिंग अधिकारी	Address/पता	Quantity/मात्रा	Delivery Days/डिलीवरी के दिन
1	Kanchan Bala	226031,CSIR-Central Drug Research Institute Sector 10, Jankipuram Extension, Sitapur Road,Lucknow 226031, Uttar Pradesh, INDIA	1	45

## Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

### 1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 25 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity by up to 25% of the contracted quantity during the currency of the contract at the contracted rates. Bidders are bound to accept the orders accordingly.

### 2. Generic

**Manufacturer Authorization:**Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid

### 3. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of

The Director CDRI, Lucknow  
payable at  
Lucknow

. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

### 4. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of

The Director, CDRI, Lucknow  
A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

### 5. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

### 6. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Sr. No.	Main component	Detailed general specifications
1.	Gas Chromatograph	<ul style="list-style-type: none"> <li>i. Gas chromatograph for analysis of Liquid Samples</li> <li>ii. GC with Flame Ionization along with licensed window-based software and two or more capillary columns (polar &amp; non-polar or equivalent, 60 m x 0.25 mm I.D) for the analysis.</li> <li>iii. GC must be capable to accommodate three or more detectors.</li> <li>iv. GC should have touch screen display.</li> <li>v. All temperature and time functions should be micro-processor controlled and displayed on the screen/Monitor.</li> <li>vi. System must have the facility of retention time locking or equivalent feature.</li> </ul>
2.	GC Oven	<ul style="list-style-type: none"> <li>i. Operating temperature range of up to 400 ° or more.</li> <li>ii. Temperature programming ramps: 20 or more.</li> <li>iii. Temperature set point resolution 0.1 degree C or better.</li> <li>iv. Temperature program: 7 or more ramps &amp; plateaus.</li> <li>v. Cool down time from 300 °C to 50 °C is &lt; 6 minutes or better.</li> </ul>
3.	Pneumatics	System must have digital/ pneumatic/ electronic flow control (EFC/ PPC/ AFC) for all the injectors & detectors.
4.	Capillary Inlet	<ul style="list-style-type: none"> <li>i. Split/split less capillary port</li> <li>ii. Temperature: 400 °C or more</li> <li>iii. Fully EPC</li> <li>iv. Split ratio: 7500:1 or better</li> <li>v. Pressure setting range 0-100 psi or better</li> </ul>
5.	Flame Ionization detector (FID)	<ul style="list-style-type: none"> <li>i. Electronic pressure and flow control of all gases</li> <li>ii. Maximum temperature: 400 °C or more</li> <li>iii. Minimum detectable level: less than 3.0 pgC/sec.</li> <li>iv. Data acquisition rate up to 300 Hz or better</li> <li>v. Linear dynamic range: 10<sup>7</sup></li> </ul>
6.	Auto-sampler	Autosampler with minimum 10 vials capacity must be supplied.
7.	Utilities/accessories	<ul style="list-style-type: none"> <li>i. Filled gas cylinders, Regulators and Gas Panel</li> <li>ii. Compatible 5 KVA online UPS with 30 minutes backup</li> <li>iii. Compatible PC (i5, 12<sup>th</sup> generation processor, Win 10 pro licensed, 1 TB HDD, 8 GB RAM or better) and 24-inch Monitor along with Printer</li> <li>iv. Consumable items: ferrules (20 Nos.), septa (200 Nos.), liners (10 Nos.), O-rings (100 Nos.), nuts (20 Nos.), vials with cap (200 Nos.) and 10 mL syringe (at least 1 No.)</li> </ul>
8.	Software	Original window-based software with license
9.	Warranty	At least for Two years
10.	Delivery	Till Destination
11.	Installation/ Training	Should be done by Engineer without any additional costing.

12	Supply experience	Must have installed such instrument(s) in Government organizations
----	-------------------	--

Note:-

1. Bidder must submit Bid Securing Declaration as per attached format in Bid document (if not submitted along with bid, the bid will be summarily rejected)
2. The bidder/manufacture must have sold at least three quantity of equipment of equivalent specification. Bidder should attach PO and successful installation certificate along with the bid. MSEs/Startups will be exempted from past experience as per rules.
3. The bidder has to give a confirmation that the spares for quoted item will be available for at least ten years.
4. The brochure and details of quoted equipment should be available on their website.

#### 7. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

#### 8. Service & Support

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

#### 9. Warranty

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope

2 Years from the date of Installation Commissioning and testing  
(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

### Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any

Category item bunched with it.

4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

**---Thank You/धन्यवाद---**